



The Natural Medicine Company

Burgage, Blessington, Co. Wicklow.

Tel.: (045) 865 575 Fax: (045) 865 827

Website: www.naturalmedicine.ie

URGENT - FIELD SAFETY NOTICE – MEDICAL DEVICE RECALL

COMVITA™ Medihoney® Wound Gel

Manufacturer: DERMA SCIENCES, Inc. 104 Shorting Rd. Scarborough, Ontario M1S 3S4

EC Representative: INTEGRA LIFESCIENCES SERVICES (France) SAS - Immeuble Séquoia 2 - 97 Allée Alexandre Borodine - 69800 SAINT PRIEST, France - SRN : FR-AR-000002474

Primary Clinical Purpose: Assist in healing minor burns, cuts, grazes and ulcers

Impacted Products:

MEDIHONEY Antibacterial Wound Gel – 25g tube

MEDIHONEY Antibacterial Wound Gel – 50g tube

Date: 11/03/2026

Dear Valued Customer,

The purpose of this letter is to inform you that Integra LifeSciences is voluntarily recalling COMVITA™ MediHoney® Wound Gel products listed in Table 1.

Reason for Recall: Packaging failures were identified related to the MediHoney® Wound and Burn products, which could lead to a breach in the sterile barrier.

Table 1: Impacted Product Information

Product	NMC Code	Comvita Code	CE Certificate	Batch Numbers
Comvita MediHoney Antibacterial Wound Gel 25G		103435	COM389	2326, 2328, 2340, 2341, 2345, 2419, 2429, 2433
Comvita MediHoney Antibacterial Wound Gel 50G		103486	COM60060	2326, 2332, 2346

Risk To Health

If the packaging of a sterile product is inadequate and the product is used on an open wound, there is a potential risk of infection. Additionally, the inability to use the device due to packaging failures may cause inconvenience to the user and delay care. There is no long-term health consequences expected due to this problem.

No incident has been reported in Europe.